

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061721\
 Data File : VY005164.D
 Acq On : 17 Jun 2021 18:06
 Operator : SY/MD
 Sample : M2748-10
 Misc : 5.71G/5ML/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-9A

Quant Time: Jun 18 03:04:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 16 04:40:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	153657	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	278756	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	191112	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	69818	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	100856	67.884	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.760%
35) Dibromofluoromethane	7.728	113	85925	71.856	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	143.720%
50) Toluene-d8	10.185	98	344356	57.985	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	115.960%
62) 4-Bromofluorobenzene	12.483	95	54680	27.377	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	54.760%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.729	84	20718	6.554	ug/l	93
64) Tetrachloroethene	10.727	164	2329	1.366	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061721\
Data File : VY005164.D
Acq On : 17 Jun 2021 18:06
Operator : SY/MD
Sample : M2748-10
Misc : 5.71G/5ML/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-9A

Quant Time: Jun 18 03:04:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061521S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 16 04:40:35 2021
Response via : Initial Calibration

