

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061721\
 Data File : VY005165.D
 Acq On : 17 Jun 2021 18:29
 Operator : SY/MD
 Sample : M2748-12
 Misc : 5.21G/5ML/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-10A

Quant Time: Jun 18 03:04:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 16 04:40:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	152627	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	279658	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	273163	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	101834	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	100559	68.141	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	136.280%
35) Dibromofluoromethane	7.728	113	86765	72.368	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	144.740%
50) Toluene-d8	10.185	98	342025	57.407	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.820%
62) 4-Bromofluorobenzene	12.483	95	120353	60.063	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	120.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061721\
Data File : VY005165.D
Acq On : 17 Jun 2021 18:29
Operator : SY/MD
Sample : M2748-12
Misc : 5.21G/5ML/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-10A

Quant Time: Jun 18 03:04:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061521S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 16 04:40:35 2021
Response via : Initial Calibration

