

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061722\
 Data File : VY009220.D
 Acq On : 17 Jun 2022 12:34
 Operator : KP/MD
 Sample : N3348-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-5-061422

Quant Time: Jun 17 14:19:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	193401	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	336758	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	308540	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	142102	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	111124	55.005	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.000%
35) Dibromofluoromethane	7.716	113	104197	49.351	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.700%
50) Toluene-d8	10.179	98	396885	49.381	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.760%
62) 4-Bromofluorobenzene	12.477	95	134895	47.231	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.460%
Target Compounds						
65) Chlorobenzene	11.514	112	14258	2.133	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061722\
Data File : VY009220.D
Acq On : 17 Jun 2022 12:34
Operator : KP/MD
Sample : N3348-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-5-061422

Quant Time: Jun 17 14:19:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 16 02:54:06 2022
Response via : Initial Calibration

