

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061722\
 Data File : VY009226.D
 Acq On : 17 Jun 2022 14:54
 Operator : KP/MD
 Sample : N3383-04
 Misc : 7.11g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-01-(10-11.5)

Quant Time: Jun 17 22:53:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	182589	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	323723	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	300421	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	141842	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	112568	59.019	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.040%
35) Dibromofluoromethane	7.710	113	101222	49.872	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.740%
50) Toluene-d8	10.179	98	382942	49.565	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.120%
62) 4-Bromofluorobenzene	12.477	95	138086	50.295	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.600%
Target Compounds						
					Qvalue	
65) Chlorobenzene	11.514	112	13766	2.115	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061722\
Data File : VY009226.D
Acq On : 17 Jun 2022 14:54
Operator : KP/MD
Sample : N3383-04
Misc : 7.11g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-01-(10-11.5)

Quant Time: Jun 17 22:53:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 16 02:54:06 2022
Response via : Initial Calibration

