

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061722\
 Data File : VY009227.D
 Acq On : 17 Jun 2022 15:18
 Operator : KP/MD
 Sample : N3383-05
 Misc : 7.26g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-02-(15-16)

Quant Time: Jun 17 22:54:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	194190	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	338211	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	315235	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	128481	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	119168	58.747	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 117.500%			
35) Dibromofluoromethane	7.710	113	101814	48.015	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 96.040%			
50) Toluene-d8	10.173	98	411252	50.949	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 101.900%			
62) 4-Bromofluorobenzene	12.477	95	138052	48.129	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 96.260%			
Target Compounds						
16) Acetone	3.936	43	9983	19.537	ug/l	92
17) Carbon Disulfide	4.174	76	10654	1.992	ug/l	98
39) Methylcyclohexane	9.179	83	5289	1.345	ug/l	90
64) Tetrachloroethene	10.715	164	12669	4.338	ug/l	96
65) Chlorobenzene	11.514	112	14304	2.095	ug/l	99
67) Ethyl Benzene	11.587	91	15707	1.318	ug/l	97
68) m/p-Xylenes	11.691	106	12072	2.570	ug/l	91
78) n-propylbenzene	12.666	91	14194	1.213	ug/l	95
80) 1,3,5-Trimethylbenzene	12.813	105	24372	3.015	ug/l	94
84) 1,2,4-Trimethylbenzene	13.117	105	57395	7.235	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061722\
Data File : VY009227.D
Acq On : 17 Jun 2022 15:18
Operator : KP/MD
Sample : N3383-05
Misc : 7.26g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-02-(15-16)

Quant Time: Jun 17 22:54:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 16 02:54:06 2022
Response via : Initial Calibration

