

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061722\
 Data File : VY009231.D
 Acq On : 17 Jun 2022 16:51
 Operator : KP/MD
 Sample : N3373-13
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DS-7WC

Quant Time: Jun 17 22:55:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	213615	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	361659	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	338037	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	152756	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	121118	54.278	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.560%	
35) Dibromofluoromethane	7.710	113	111939	49.367	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 98.740%	
50) Toluene-d8	10.173	98	431984	50.048	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.100%	
62) 4-Bromofluorobenzene	12.477	95	151439	49.373	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 98.740%	
Target Compounds						
65) Chlorobenzene	11.508	112	15964	2.180	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061722\
Data File : VY009231.D
Acq On : 17 Jun 2022 16:51
Operator : KP/MD
Sample : N3373-13
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DS-7WC

Quant Time: Jun 17 22:55:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 16 02:54:06 2022
Response via : Initial Calibration

