

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061722\
 Data File : VY009232.D
 Acq On : 17 Jun 2022 17:15
 Operator : KP/MD
 Sample : N3371-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-211

Quant Time: Jun 17 22:55:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	200241	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	349691	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	325420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	150326	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	122192	58.417	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.840%
35) Dibromofluoromethane	7.710	113	107114	48.856	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.720%
50) Toluene-d8	10.173	98	412138	49.383	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.760%
62) 4-Bromofluorobenzene	12.477	95	142024	47.888	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.780%
Target Compounds						
					Qvalue	
65) Chlorobenzene	11.514	112	15126	2.146	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061722\
Data File : VY009232.D
Acq On : 17 Jun 2022 17:15
Operator : KP/MD
Sample : N3371-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-211

Quant Time: Jun 17 22:55:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 16 02:54:06 2022
Response via : Initial Calibration

