

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061722\
 Data File : VY009235.D
 Acq On : 17 Jun 2022 18:23
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 17 22:57:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	0.378	0.362	4.2	111	0.00
3 P	Chloromethane	0.469	0.387	17.5	92	0.00
4 C	Vinyl Chloride	0.492	0.430	12.6#	93	0.00
5 T	Bromomethane	0.346	0.313	9.5	101	0.00
6 T	Chloroethane	0.308	0.288	6.5	98	0.00
7 T	Trichlorofluoromethane	0.786	0.800	-1.8	113	0.00
8 T	Diethyl Ether	0.298	0.306	-2.7	109	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.483	0.503	-4.1	112	0.00
10 T	Methyl Iodide	0.618	0.660	-6.8	107	0.00
11 T	Tert butyl alcohol	0.083	0.055	33.7#	101	0.00
12 CM	1,1-Dichloroethene	0.476	0.476	0.0	107	0.00
13 T	Acrolein	0.017	0.010	41.2#	82	-0.01
14 T	Allyl chloride	0.859	0.843	1.9	101	0.00
15 T	Acrylonitrile	0.160	0.162	-1.3	105	0.00
16 T	Acetone	0.132	0.123	6.8	107	0.00
17 T	Carbon Disulfide	1.377	1.378	-0.1	105	0.00
18 T	Methyl Acetate	0.437	0.378	13.5	107	0.00
19 T	Methyl tert-butyl Ether	1.436	1.495	-4.1	108	0.00
20 T	Methylene Chloride	0.670	0.631	5.8	113	0.00
21 T	trans-1,2-Dichloroethene	0.557	0.578	-3.8	104	-0.01
22 T	Diisopropyl ether	1.770	1.732	2.1	100	0.00
23 T	Vinyl Acetate	1.114	1.141	-2.4	100	0.00
24 P	1,1-Dichloroethane	1.000	1.020	-2.0	105	0.00
25 T	2-Butanone	0.208	0.203	2.4	102	0.00
26 T	2,2-Dichloropropane	0.886	0.858	3.2	103	0.00
27 T	cis-1,2-Dichloroethene	0.654	0.674	-3.1	105	0.00
28 T	Bromochloromethane	0.429	0.419	2.3	99	0.00
29 T	Tetrahydrofuran	0.138	0.135	2.2	99	0.00
30 C	Chloroform	1.029	1.051	-2.1#	106	0.00
31 T	Cyclohexane	0.936	0.881	5.9	99	0.00
32 T	1,1,1-Trichloroethane	0.905	0.918	-1.4	104	0.00
33 S	1,2-Dichloroethane-d4	0.522	0.532	-1.9	108	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
35 S	Dibromofluoromethane	0.313	0.328	-4.8	107	0.00
36 T	1,1-Dichloropropene	0.482	0.479	0.6	99	0.00
37 T	Ethyl Acetate	0.287	0.284	1.0	102	0.00
38 T	Carbon Tetrachloride	0.499	0.506	-1.4	103	0.00
39 T	Methylcyclohexane	0.581	0.599	-3.1	102	0.00
40 TM	Benzene	1.408	1.405	0.2	101	0.00
41 T	Methacrylonitrile	0.160	0.163	-1.9	93	-0.01
42 TM	1,2-Dichloroethane	0.411	0.415	-1.0	105	0.00
43 T	Isopropyl Acetate	0.531	0.527	0.8	102	0.00
44 TM	Trichloroethene	0.403	0.411	-2.0	105	0.00
45 C	1,2-Dichloropropane	0.357	0.361	-1.1#	102	0.00
46 T	Dibromomethane	0.205	0.209	-2.0	103	0.00
47 T	Bromodichloromethane	0.489	0.501	-2.5	104	0.00
48 T	Methyl methacrylate	0.237	0.237	0.0	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061722\
 Data File : VY009235.D
 Acq On : 17 Jun 2022 18:23
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 17 22:57:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 16 02:54:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	94	0.00
50 S	Toluene-d8	1.193	1.186	0.6	103	0.00
51 T	4-Methyl-2-Pentanone	0.279	0.272	2.5	101	0.00
52 CM	Toluene	0.899	0.882	1.9#	101	0.00
53 T	t-1,3-Dichloropropene	0.513	0.531	-3.5	105	0.00
54 T	cis-1,3-Dichloropropene	0.580	0.591	-1.9	103	0.00
55 T	1,1,2-Trichloroethane	0.291	0.299	-2.7	103	0.00
56 T	Ethyl methacrylate	0.394	0.411	-4.3	103	0.00
57 T	1,3-Dichloropropane	0.496	0.510	-2.8	104	0.00
58 T	2-Chloroethyl Vinyl ether	0.171	0.170	0.6	106	0.00
59 T	2-Hexanone	0.192	0.187	2.6	102	0.00
60 T	Dibromochloromethane	0.356	0.374	-5.1	107	0.00
61 T	1,2-Dibromoethane	0.283	0.289	-2.1	102	0.00
62 S	4-Bromofluorobenzene	0.424	0.422	0.5	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	106	0.00
64 T	Tetrachloroethene	0.463	0.480	-3.7	105	0.00
65 PM	Chlorobenzene	1.083	1.075	0.7	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.397	0.409	-3.0	105	0.00
67 C	Ethyl Benzene	1.890	1.870	1.1#	101	0.00
68 T	m/p-Xylenes	0.745	0.735	1.3	101	0.00
69 T	o-Xylene	0.706	0.722	-2.3	104	0.00
70 T	Styrene	1.184	1.217	-2.8	105	0.00
71 P	Bromoform	0.253	0.261	-3.2	106	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
73 T	Isopropylbenzene	3.763	3.738	0.7	103	0.00
74 T	N-amyl acetate	1.050	1.011	3.7	103	0.00
75 P	1,1,2,2-Tetrachloroethane	0.745	0.700	6.0	100	0.00
76 T	1,2,3-Trichloropropane	0.552	0.501	9.2	99	0.00
77 T	Bromobenzene	0.925	0.930	-0.5	106	0.00
78 T	n-propylbenzene	4.553	4.451	2.2	102	0.00
79 T	2-Chlorotoluene	2.565	2.509	2.2	103	0.00
80 T	1,3,5-Trimethylbenzene	3.145	3.092	1.7	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.277	0.261	5.8	101	0.00
82 T	4-Chlorotoluene	2.677	2.594	3.1	103	0.00
83 T	tert-Butylbenzene	2.765	2.770	-0.2	102	0.00
84 T	1,2,4-Trimethylbenzene	3.087	3.079	0.3	103	0.00
85 T	sec-Butylbenzene	4.095	4.049	1.1	103	0.00
86 T	p-Isopropyltoluene	3.398	3.371	0.8	103	0.00
87 T	1,3-Dichlorobenzene	1.783	1.777	0.3	106	0.00
88 T	1,4-Dichlorobenzene	1.810	1.759	2.8	103	0.00
89 T	n-Butylbenzene	3.213	3.159	1.7	103	0.00
90 T	Hexachloroethane	0.641	0.634	1.1	104	0.00
91 T	1,2-Dichlorobenzene	1.613	1.608	0.3	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.134	0.128	4.5	98	0.00
93 T	1,2,4-Trichlorobenzene	1.014	0.959	5.4	96	0.00
94 T	Hexachlorobutadiene	0.557	0.526	5.6	98	0.00
95 T	Naphthalene	2.115	2.131	-0.8	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061722\
Data File : VY009235.D
Acq On : 17 Jun 2022 18:23
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jun 17 22:57:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061522S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 16 02:54:06 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.892	0.841	5.7	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 5