

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061824\
 Data File : VY018631.D
 Acq On : 18 Jun 2024 15:56
 Operator : SY/MD
 Sample : P2923-02
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOC

Quant Time: Jun 19 02:02:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 14 07:08:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	132467	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	259483	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	244052	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	90995	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	77946	58.445	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.880%
35) Dibromofluoromethane	7.704	113	79957	50.513	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.020%
50) Toluene-d8	10.173	98	316780	52.040	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	104.080%
62) 4-Bromofluorobenzene	12.477	95	90353	47.005	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061824\
Data File : VY018631.D
Acq On : 18 Jun 2024 15:56
Operator : SY/MD
Sample : P2923-02
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC

Quant Time: Jun 19 02:02:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 14 07:08:24 2024
Response via : Initial Calibration

