

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061824\
 Data File : VY018622.D
 Acq On : 18 Jun 2024 10:09
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 19 01:58:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 14 07:08:24 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.329	0.275	16.4	76	0.00
3 P	Chloromethane	0.678	0.734	-8.3	89	0.00
4 C	Vinyl Chloride	0.840	0.912	-8.6#	92	0.00
5 T	Bromomethane	0.624	0.579	7.2	83	0.00
6 T	Chloroethane	0.550	0.544	1.1	85	0.00
7 T	Trichlorofluoromethane	0.904	0.955	-5.6	91	0.00
8 T	Diethyl Ether	0.258	0.235	8.9	75	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.473	0.453	4.2	84	0.00
10 T	Methyl Iodide	0.515	0.502	2.5	73	0.00
11 T	Tert butyl alcohol	0.091	0.032	64.8#	48#	0.00
12 CM	1,1-Dichloroethene	0.475	0.440	7.4#	79	0.00
13 T	Acrolein	0.029	0.025	13.8	66	0.00
14 T	Allyl chloride	0.640	0.648	-1.3	85	0.00
15 T	Acrylonitrile	0.115	0.106	7.8	76	0.00
16 T	Acetone	0.102	0.110	-7.8	87	0.00
17 T	Carbon Disulfide	1.266	1.275	-0.7	77	0.00
18 T	Methyl Acetate	0.260	0.244	6.2	71	0.00
19 T	Methyl tert-butyl Ether	1.267	1.140	10.0	76	0.00
20 T	Methylene Chloride	0.854	0.534	37.5#	75	0.00
21 T	trans-1,2-Dichloroethene	0.541	0.513	5.2	81	0.00
22 T	Diisopropyl ether	1.445	1.492	-3.3	86	0.00
23 T	Vinyl Acetate	1.042	1.188	-14.0	82	0.00
24 P	1,1-Dichloroethane	0.931	0.885	4.9	82	0.00
25 T	2-Butanone	0.154	0.155	-0.6	81	0.00
26 T	2,2-Dichloropropane	0.785	0.750	4.5	85	0.00
27 T	cis-1,2-Dichloroethene	0.653	0.622	4.7	82	0.00
28 T	Bromochloromethane	0.382	0.326	14.7	104	0.00
29 T	Tetrahydrofuran	0.094	0.089	5.3	76	0.00
30 C	Chloroform	1.015	0.967	4.7#	82	0.00
31 T	Cyclohexane	0.771	0.773	-0.3	90	0.00
32 T	1,1,1-Trichloroethane	0.862	0.828	3.9	83	0.00
33 S	1,2-Dichloroethane-d4	0.503	0.487	3.2	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.305	0.300	1.6	96	0.00
36 T	1,1-Dichloropropene	0.420	0.413	1.7	84	0.00
37 T	Ethyl Acetate	0.198	0.192	3.0	79	0.00
38 T	Carbon Tetrachloride	0.440	0.433	1.6	84	0.00
39 T	Methylcyclohexane	0.514	0.542	-5.4	86	0.00
40 TM	Benzene	1.298	1.296	0.2	83	0.00
41 T	Methacrylonitrile	0.100	0.101	-1.0	84	-0.01
42 TM	1,2-Dichloroethane	0.339	0.321	5.3	79	0.00
43 T	Isopropyl Acetate	0.369	0.365	1.1	79	0.00
44 TM	Trichloroethene	0.340	0.322	5.3	80	0.00
45 C	1,2-Dichloropropane	0.297	0.296	0.3#	84	0.00
46 T	Dibromomethane	0.184	0.174	5.4	78	0.00
47 T	Bromodichloromethane	0.440	0.433	1.6	81	0.00
48 T	Methyl methacrylate	0.166	0.167	-0.6	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061824\
 Data File : VY018622.D
 Acq On : 18 Jun 2024 10:09
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 19 01:58:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 14 07:08:24 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	73	0.00
50 S	Toluene-d8	1.173	1.162	0.9	97	0.00
51 T	4-Methyl-2-Pentanone	0.195	0.192	1.5	77	0.00
52 CM	Toluene	0.834	0.847	-1.6#	83	0.00
53 T	t-1,3-Dichloropropene	0.384	0.394	-2.6	81	0.00
54 T	cis-1,3-Dichloropropene	0.464	0.471	-1.5	83	0.00
55 T	1,1,2-Trichloroethane	0.245	0.233	4.9	80	0.00
56 T	Ethyl methacrylate	0.321	0.321	0.0	76	0.00
57 T	1,3-Dichloropropane	0.407	0.395	2.9	80	0.00
58 T	2-Chloroethyl Vinyl ether	0.155	0.147	5.2	84	0.00
59 T	2-Hexanone	0.136	0.143	-5.1	81	0.00
60 T	Dibromochloromethane	0.304	0.297	2.3	79	0.00
61 T	1,2-Dibromoethane	0.226	0.217	4.0	76	0.00
62 S	4-Bromofluorobenzene	0.370	0.384	-3.8	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.357	0.352	1.4	82	0.00
65 PM	Chlorobenzene	1.033	1.031	0.2	82	0.00
66 T	1,1,1,2-Tetrachloroethane	0.346	0.341	1.4	81	0.00
67 C	Ethyl Benzene	1.792	1.816	-1.3#	82	0.00
68 T	m/p-Xylenes	0.697	0.724	-3.9	82	0.00
69 T	o-Xylene	0.671	0.674	-0.4	81	0.00
70 T	Styrene	1.108	1.171	-5.7	83	0.00
71 P	Bromoform	0.190	0.185	2.6	79	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.651	3.564	2.4	83	0.00
74 T	N-amyl acetate	0.748	0.763	-2.0	79	0.00
75 P	1,1,2,2-Tetrachloroethane	0.648	0.600	7.4	76	0.00
76 T	1,2,3-Trichloropropane	0.469	0.456	2.8	76	0.00
77 T	Bromobenzene	0.844	0.797	5.6	82	0.00
78 T	n-propylbenzene	4.273	4.281	-0.2	84	0.00
79 T	2-Chlorotoluene	2.412	2.348	2.7	81	0.00
80 T	1,3,5-Trimethylbenzene	2.894	2.908	-0.5	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.179	0.180	-0.6	75	0.00
82 T	4-Chlorotoluene	2.429	2.407	0.9	83	0.00
83 T	tert-Butylbenzene	2.678	2.583	3.5	83	0.00
84 T	1,2,4-Trimethylbenzene	2.892	2.934	-1.5	84	0.00
85 T	sec-Butylbenzene	3.820	3.903	-2.2	85	0.00
86 T	p-Isopropyltoluene	3.238	3.296	-1.8	85	0.00
87 T	1,3-Dichlorobenzene	1.703	1.661	2.5	83	0.00
88 T	1,4-Dichlorobenzene	1.671	1.604	4.0	82	0.00
89 T	n-Butylbenzene	2.919	3.063	-4.9	87	0.00
90 T	Hexachloroethane	0.588	0.630	-7.1	87	0.00
91 T	1,2-Dichlorobenzene	1.455	1.420	2.4	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.094	0.083	11.7	72	0.00
93 T	1,2,4-Trichlorobenzene	0.811	0.757	6.7	79	0.00
94 T	Hexachlorobutadiene	0.416	0.412	1.0	85	0.00
95 T	Naphthalene	1.515	1.445	4.6	75	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061824\
Data File : VY018622.D
Acq On : 18 Jun 2024 10:09
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jun 19 01:58:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 14 07:08:24 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.685	0.645	5.8	80	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6