

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061923\
 Data File : VY014228.D
 Acq On : 19 Jun 2023 14:34
 Operator : KP/MD
 Sample : VY0619SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0619SBL01

Quant Time: Jun 19 17:02:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	229717	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	322279	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	304364	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	170926	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	131502	41.109	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	82.220%
35) Dibromofluoromethane	7.710	113	110628	48.404	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.800%
50) Toluene-d8	10.173	98	367257	42.981	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	85.960%
62) 4-Bromofluorobenzene	12.477	95	140908	46.423	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061923\
Data File : VY014228.D
Acq On : 19 Jun 2023 14:34
Operator : KP/MD
Sample : VY0619SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0619SBL01

Quant Time: Jun 19 17:02:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

