

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061923\
 Data File : VY014235.D
 Acq On : 19 Jun 2023 18:08
 Operator : KP/MD
 Sample : 03295-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1

Quant Time: Jun 20 04:59:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	2282	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	4533	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	5794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	2867	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	3684	115.932	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	231.860%#
35) Dibromofluoromethane	7.716	113	1386	43.114	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.220%
50) Toluene-d8	10.179	98	5706	47.477	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.960%
62) 4-Bromofluorobenzene	12.477	95	2253	52.772	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	105.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061923\
Data File : VY014235.D
Acq On : 19 Jun 2023 18:08
Operator : KP/MD
Sample : 03295-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1

Quant Time: Jun 20 04:59:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

