

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061923\
 Data File : VY014237.D
 Acq On : 19 Jun 2023 18:55
 Operator : KP/MD
 Sample : 03255-06RE
 Misc : 6.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOTAL-VOC-SB-24-GRAB-7RE

Quant Time: Jun 20 05:00:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	43978	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	61956	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	48532	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	14513	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	22399	36.576	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	73.160%
35) Dibromofluoromethane	7.716	113	20413	46.459	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.920%
50) Toluene-d8	10.179	98	73648	44.835	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.660%
62) 4-Bromofluorobenzene	12.483	95	15313	26.242	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	52.480%
Target Compounds					Qvalue	
16) Acetone	3.948	43	2956	16.013	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061923\
Data File : VY014237.D
Acq On : 19 Jun 2023 18:55
Operator : KP/MD
Sample : 03255-06RE
Misc : 6.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOTAL-VOC-SB-24-GRAB-7RE

Quant Time: Jun 20 05:00:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

