

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062023\
 Data File : VY014275.D
 Acq On : 20 Jun 2023 12:01
 Operator : KP/MD
 Sample : VY0620SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0620SBL01

Quant Time: Jun 21 02:06:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	187103	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	297055	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	276820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	130985	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	126063	48.385	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.760%
35) Dibromofluoromethane	7.716	113	97446	46.256	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.520%
50) Toluene-d8	10.179	98	355610	45.152	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.300%
62) 4-Bromofluorobenzene	12.483	95	111942	40.011	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062023\
Data File : VY014275.D
Acq On : 20 Jun 2023 12:01
Operator : KP/MD
Sample : VY0620SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0620SBL01

Quant Time: Jun 21 02:06:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

