

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062023\
 Data File : VY014305.D
 Acq On : 21 Jun 2023 00:52
 Operator : KP/MD
 Sample : 03321-07
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-23

Quant Time: Jun 21 02:12:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	188675	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.697	114	302673	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	280873	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	133482	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	128310	48.837	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.680%
35) Dibromofluoromethane	7.728	113	99196	46.213	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.420%
50) Toluene-d8	10.185	98	361043	44.991	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.980%
62) 4-Bromofluorobenzene	12.483	95	117906	41.361	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.720%
Target Compounds						
20) Methylene Chloride	4.729	84	24658	6.644	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062023\
Data File : VY014305.D
Acq On : 21 Jun 2023 00:52
Operator : KP/MD
Sample : 03321-07
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-23

Quant Time: Jun 21 02:12:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

