

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062023\
 Data File : VY014306.D
 Acq On : 21 Jun 2023 01:16
 Operator : KP/MD
 Sample : 03321-10
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-22

Quant Time: Jun 21 02:12:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	89363	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.697	114	145969	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	148914	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	68395	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	74104	59.550	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.100%
35) Dibromofluoromethane	7.728	113	53564	51.744	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.480%
50) Toluene-d8	10.185	98	177040	45.745	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.500%
62) 4-Bromofluorobenzene	12.489	95	59398	43.205	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.420%
Target Compounds						
20) Methylene Chloride	4.728	84	11500	6.483	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062023\
Data File : VY014306.D
Acq On : 21 Jun 2023 01:16
Operator : KP/MD
Sample : 03321-10
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-22

Quant Time: Jun 21 02:12:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

