

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY062220\
 Data File : VY002967.D
 Acq On : 22 Jun 2020 12:35
 Operator : SY/MD
 Sample : L3024-16RE
 Misc : 5.03G/5ML/MSVOA Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RO-6RE

Quant Time: Jun 22 14:37:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y061920S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 20 06:39:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	222704	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.69	114	359406	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	344565	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.42	152	175301	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.15	65	113255	61.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.14%	
35) Dibromofluoromethane	7.72	113	110440	54.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.62%	
50) Toluene-d8	10.18	98	422245	52.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.32%	
62) 4-Bromofluorobenzene	12.48	95	146842	52.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.58%	
Target Compounds						
3) Chloromethane	2.11	50	4664	2.893	ug/l	99
20) Methylene Chloride	4.73	84	16633	1.046	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY062220\
Data File : VY002967.D
Acq On : 22 Jun 2020 12:35
Operator : SY/MD
Sample : L3024-16RE
Misc : 5.03G/5ML/MSVOA Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RO-6RE

Quant Time: Jun 22 14:37:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y061920S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 20 06:39:15 2020
Response via : Initial Calibration

