



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062221\
 Data File : VY005190.D
 Acq On : 22 Jun 2021 11:13
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
 Title : SW846 8260
 Last Update : Wed Jun 23 07:57:52 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	24077	PASS
75	95	30	60	49.2	68293	PASS
95	95	100	100	100.0	138701	PASS
96	95	5	9	6.7	9281	PASS
173	174	0.00	2	1.0	1332	PASS
174	95	50	100	94.4	130952	PASS
175	174	5	9	7.6	10004	PASS
176	174	95	101	97.5	127645	PASS
177	176	5	9	6.8	8694	PASS