

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062322\
 Data File : VY009300.D
 Acq On : 23 Jun 2022 17:31
 Operator : KP/MD
 Sample : N3462-05
 Misc : 4.50g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-6-4-6-20220622

Quant Time: Jun 24 09:24:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 21 14:08:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	184969	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	329277	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	303431	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	133471	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	113596	55.919	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.840%
35) Dibromofluoromethane	7.716	113	100996	47.361	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.720%
50) Toluene-d8	10.179	98	386535	47.185	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.380%
62) 4-Bromofluorobenzene	12.477	95	127502	44.294	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	88.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062322\
Data File : VY009300.D
Acq On : 23 Jun 2022 17:31
Operator : KP/MD
Sample : N3462-05
Misc : 4.50g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-6-4-6-20220622

Quant Time: Jun 24 09:24:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062122S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 21 14:08:27 2022
Response via : Initial Calibration

