

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062322\
 Data File : VY009301.D
 Acq On : 23 Jun 2022 17:54
 Operator : KP/MD
 Sample : N3462-06
 Misc : 4.20g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-6-18-20-20220622

Quant Time: Jun 24 09:24:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 21 14:08:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	181119	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	316875	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	302203	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	140119	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	108329	54.459	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.920%
35) Dibromofluoromethane	7.716	113	96817	47.178	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.360%
50) Toluene-d8	10.179	98	383263	48.617	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.240%
62) 4-Bromofluorobenzene	12.477	95	129155	46.624	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062322\
Data File : VY009301.D
Acq On : 23 Jun 2022 17:54
Operator : KP/MD
Sample : N3462-06
Misc : 4.20g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-6-18-20-20220622

Quant Time: Jun 24 09:24:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062122S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 21 14:08:27 2022
Response via : Initial Calibration

