

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062322\
 Data File : VY009303.D
 Acq On : 23 Jun 2022 18:41
 Operator : KP/MD
 Sample : N3462-08
 Misc : 2.74g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-2-3-5-20220622

Quant Time: Jun 24 09:24:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 21 14:08:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	180479	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	309775	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	295888	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	132594	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	104288	52.614	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.220%
35) Dibromofluoromethane	7.716	113	92326	46.021	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.040%
50) Toluene-d8	10.179	98	375049	48.666	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.340%
62) 4-Bromofluorobenzene	12.477	95	124897	46.121	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062322\
Data File : VY009303.D
Acq On : 23 Jun 2022 18:41
Operator : KP/MD
Sample : N3462-08
Misc : 2.74g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-2-3-5-20220622

Quant Time: Jun 24 09:24:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062122S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 21 14:08:27 2022
Response via : Initial Calibration

