

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062322\
 Data File : VY009302.D
 Acq On : 23 Jun 2022 18:18
 Operator : KP/MD
 Sample : N3462-07
 Misc : 3.61g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-2-0-2-20220622

Quant Time: Jun 24 09:24:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 21 14:08:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	176381	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	312985	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	293423	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	129354	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	106883	55.176	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.360%
35) Dibromofluoromethane	7.710	113	95836	47.281	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.560%
50) Toluene-d8	10.179	98	371778	47.746	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.500%
62) 4-Bromofluorobenzene	12.477	95	122561	44.794	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.580%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062322\
Data File : VY009302.D
Acq On : 23 Jun 2022 18:18
Operator : KP/MD
Sample : N3462-07
Misc : 3.61g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-2-0-2-20220622

Quant Time: Jun 24 09:24:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062122S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 21 14:08:27 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY009302.D
Acq: 23 Jun 2022 18:18




Instrument :
MSVOA_Y
ClientSampleId :
SB-2-0-2-20220622


Tgt Ion:168 Resp: 176381
Ion Ratio Lower Upper
168 100
99 52.9 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 55.176 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY009302.D
Acq: 23 Jun 2022 18:18

Tgt Ion: 65 Resp: 106883
Ion Ratio Lower Upper
65 100
67 54.3 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY009302.D

Acq: 23 Jun 2022 18:18

Instrument :

MSVOA_Y

ClientSampleId :

SB-2-0-2-20220622





Tgt Ion:114 Resp: 312985

Ion Ratio Lower Upper

114 100

63 21.0 0.0 43.2

88 15.1 0.0 31.8



#35

Dibromofluoromethane

Concen: 47.281 ug/l

RT: 7.710 min Scan# 966

Delta R.T. -0.006 min

Lab File: VY009302.D

Acq: 23 Jun 2022 18:18




Tgt Ion:113 Resp: 95836


Ion Ratio Lower Upper

113 100

111 104.5 82.6 124.0

192 19.9 17.6 26.4









#50
Toluene-d8
Concen: 47.746 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY009302.D
Acq: 23 Jun 2022 18:18




Instrument :
MSVOA_Y
ClientSampleId :
SB-2-0-2-20220622


Tgt Ion: 98 Resp: 371778
Ion Ratio Lower Upper
98 100
100 65.2 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 44.794 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. 0.000 min
Lab File: VY009302.D
Acq: 23 Jun 2022 18:18

Tgt Ion: 95 Resp: 122561
Ion Ratio Lower Upper
95 100
174 86.0 0.0 177.0
176 84.4 0.0 175.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.489 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY009302.D


Acq: 23 Jun 2022 18:18

Instrument :

MSVOA_Y

ClientSampleId :

SB-2-0-2-20220622






Tgt Ion:117 Resp: 293423

Ion Ratio Lower Upper

117 100

82 52.7 43.0 64.4

119 32.4 25.4 38.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. 0.000 min

Lab File: VY009302.D

Acq: 23 Jun 2022 18:18

Tgt Ion:152 Resp: 129354

Ion Ratio Lower Upper

152 100

115 55.8 28.2 84.7

150 157.3 0.0 347.6

