



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062421\
 Data File : VY005198.D
 Acq On : 24 Jun 2021 09:30
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
 Title : SW846 8260
 Last Update : Wed Jun 23 07:57:52 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	18422	PASS
75	95	30	60	47.0	51709	PASS
95	95	100	100	100.0	110123	PASS
96	95	5	9	6.8	7441	PASS
173	174	0.00	2	1.0	1012	PASS
174	95	50	100	96.7	106480	PASS
175	174	5	9	7.6	8077	PASS
176	174	95	101	97.4	103741	PASS
177	176	5	9	7.2	7517	PASS