


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv062520\
 Data File : VY002990.D
 Acq On : 25 Jun 2020 10:27
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y062520S.M
 Title : SW846 8260
 Last Update : Thu Jun 25 18:46:18 2020



AutoFind: Scans 1761, 1762, 1763; Background Corrected with Scan 1753

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.2	21762	PASS
75	95	30	60	51.1	61208	PASS
95	95	100	100	100.0	119770	PASS
96	95	5	9	7.0	8409	PASS
173	174	0.00	2	1.4	1639	PASS
174	95	50	100	99.1	118682	PASS
175	174	5	9	7.7	9150	PASS
176	174	95	101	97.7	115944	PASS
177	176	5	9	7.8	9091	PASS