

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062623\
 Data File : VY014434.D
 Acq On : 26 Jun 2023 15:17
 Operator : KP/MD
 Sample : 03402-05
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP3-0623

Quant Time: Jun 27 01:48:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 17:18:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	913	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.697	114	2028	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	2117	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	445	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	1308	115.388	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	230.780%#
35) Dibromofluoromethane	7.722	113	248	19.793	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	39.580%#
50) Toluene-d8	10.185	98	2325	46.681	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.360%
62) 4-Bromofluorobenzene	12.477	95	520	29.245	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	58.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062623\
Data File : VY014434.D
Acq On : 26 Jun 2023 15:17
Operator : KP/MD
Sample : 03402-05
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP3-0623

Quant Time: Jun 27 01:48:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 17:18:46 2023
Response via : Initial Calibration

