

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062624\
 Data File : VY018705.D
 Acq On : 26 Jun 2024 14:16
 Operator : SY/MD
 Sample : P3076-01
 Misc : 6.87g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 7-N-1

Quant Time: Jun 27 07:51:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 14 07:08:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	100298	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	184795	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	164998	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	64500	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	51323	50.825	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 101.660%	
35) Dibromofluoromethane	7.704	113	57606	51.102	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.200%	
50) Toluene-d8	10.173	98	209944	48.429	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 96.860%	
62) 4-Bromofluorobenzene	12.477	95	63878	46.663	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 93.320%	
Target Compounds						
					Qvalue	
64) Tetrachloroethene	10.715	164	1480	1.258	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062624\
Data File : VY018705.D
Acq On : 26 Jun 2024 14:16
Operator : SY/MD
Sample : P3076-01
Misc : 6.87g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
7-N-1

Quant Time: Jun 27 07:51:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 14 07:08:24 2024
Response via : Initial Calibration

