

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062624\
 Data File : VY018708.D
 Acq On : 26 Jun 2024 15:31
 Operator : SY/MD
 Sample : P3076-10
 Misc : 6.74g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 7-S-1

Quant Time: Jun 27 07:51:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 14 07:08:24 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	99947	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	184384	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	165630	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	64714	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	51793	51.471	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.940%	
35) Dibromofluoromethane	7.704	113	56872	50.563	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.120%	
50) Toluene-d8	10.173	98	210843	48.744	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 97.480%	
62) 4-Bromofluorobenzene	12.477	95	62182	45.526	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 91.060%	
Target Compounds						Qvalue
20) Methylene Chloride	4.698	84	20598	8.974	ug/l	94
64) Tetrachloroethene	10.709	164	1849	1.565	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062624\
Data File : VY018708.D
Acq On : 26 Jun 2024 15:31
Operator : SY/MD
Sample : P3076-10
Misc : 6.74g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
7-S-1

Quant Time: Jun 27 07:51:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 14 07:08:24 2024
Response via : Initial Calibration

