

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062722\
 Data File : VY009345.D
 Acq On : 27 Jun 2022 16:08
 Operator : KP/MD
 Sample : N3426-03
 Misc : 5.96g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B2207-27.5

Quant Time: Jun 28 04:30:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 25 07:08:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	137408	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	231345	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	210253	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	88650	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	59184	57.916	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.840%
35) Dibromofluoromethane	7.710	113	62217	49.825	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.660%
50) Toluene-d8	10.173	98	187980	51.763	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.520%
62) 4-Bromofluorobenzene	12.477	95	84677	47.614	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.220%
Target Compounds						
16) Acetone	3.936	43	6898	19.636	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062722\
Data File : VY009345.D
Acq On : 27 Jun 2022 16:08
Operator : KP/MD
Sample : N3426-03
Misc : 5.96g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B2207-27.5

Quant Time: Jun 28 04:30:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 25 07:08:46 2022
Response via : Initial Calibration

