

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062722\
 Data File : VY009351.D
 Acq On : 27 Jun 2022 18:24
 Operator : KP/MD
 Sample : N3434-06
 Misc : 6.85g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B2205-35

Quant Time: Jun 28 04:31:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 25 07:08:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	125151	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	210812	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	191141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	87865	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	52936	56.784	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.560%
35) Dibromofluoromethane	7.716	113	54657	48.034	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.060%
50) Toluene-d8	10.173	98	163384	49.009	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.020%
62) 4-Bromofluorobenzene	12.477	95	71262	43.973	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062722\
Data File : VY009351.D
Acq On : 27 Jun 2022 18:24
Operator : KP/MD
Sample : N3434-06
Misc : 6.85g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B2205-35

Quant Time: Jun 28 04:31:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 25 07:08:46 2022
Response via : Initial Calibration

