

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062722\
 Data File : VY009352.D
 Acq On : 27 Jun 2022 18:47
 Operator : KP/MD
 Sample : N3434-08
 Misc : 6.98g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B2206-35

Quant Time: Jun 28 04:32:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 25 07:08:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	118335	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	199891	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	182447	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	82890	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	50808	57.717	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.440%
35) Dibromofluoromethane	7.710	113	53942	49.996	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.000%
50) Toluene-d8	10.179	98	157425	49.929	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.860%
62) 4-Bromofluorobenzene	12.477	95	70187	45.676	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.360%
Target Compounds						
20) Methylene Chloride	4.704	84	12779	6.872	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062722\
Data File : VY009352.D
Acq On : 27 Jun 2022 18:47
Operator : KP/MD
Sample : N3434-08
Misc : 6.98g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B2206-35

Quant Time: Jun 28 04:32:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 25 07:08:46 2022
Response via : Initial Calibration

