

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062723\
 Data File : VY014456.D
 Acq On : 27 Jun 2023 14:25
 Operator : KP/MD
 Sample : 03403-01RE
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ORA-1652RE

Quant Time: Jun 28 01:56:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 17:18:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	20911	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	28789	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	24780	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	11354	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	12593	48.504	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.000%
35) Dibromofluoromethane	7.722	113	9064	50.960	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.920%
50) Toluene-d8	10.179	98	32382	45.800	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.600%
62) 4-Bromofluorobenzene	12.483	95	9225	36.547	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	73.100%
Target Compounds						
16) Acetone	3.954	43	1358	16.623	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062723\
Data File : VY014456.D
Acq On : 27 Jun 2023 14:25
Operator : KP/MD
Sample : 03403-01RE
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ORA-1652RE

Quant Time: Jun 28 01:56:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 17:18:46 2023
Response via : Initial Calibration

