

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062723\
 Data File : VY014464.D
 Acq On : 27 Jun 2023 17:32
 Operator : KP/MD
 Sample : 03392-08
 Misc : 6.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-64

Quant Time: Jun 28 02:27:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 17:18:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	159841	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	250037	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	233773	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	106810	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	113946	57.416	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.840%
35) Dibromofluoromethane	7.716	113	80123	51.867	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.740%
50) Toluene-d8	10.179	98	297419	48.434	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.860%
62) 4-Bromofluorobenzene	12.483	95	91335	41.662	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062723\
Data File : VY014464.D
Acq On : 27 Jun 2023 17:32
Operator : KP/MD
Sample : 03392-08
Misc : 6.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-64

Quant Time: Jun 28 02:27:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 17:18:46 2023
Response via : Initial Calibration

