

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062723\
 Data File : VY014466.D
 Acq On : 27 Jun 2023 18:19
 Operator : KP/MD
 Sample : 03428-01
 Misc : 6.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PNNL-APM-0621-1

Quant Time: Jun 28 02:28:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 17:18:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	182625	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	281368	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	264981	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	125862	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	122962	54.229	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.460%
35) Dibromofluoromethane	7.716	113	88174	50.723	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.440%
50) Toluene-d8	10.179	98	336367	48.677	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.360%
62) 4-Bromofluorobenzene	12.483	95	108229	43.871	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	87.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062723\
Data File : VY014466.D
Acq On : 27 Jun 2023 18:19
Operator : KP/MD
Sample : 03428-01
Misc : 6.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PNNL-APM-0621-1

Quant Time: Jun 28 02:28:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 17:18:46 2023
Response via : Initial Calibration

