



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062723\
Data File : VY014447.D
Acq On : 27 Jun 2023 09:44
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
Title : SW846 8260
Last Update : Fri Jun 23 17:18:46 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.7	14551	PASS
75	95	30	60	56.8	39904	PASS
95	95	100	100	100.0	70200	PASS
96	95	5	9	6.9	4865	PASS
173	174	0.00	2	0.7	461	PASS
174	95	50	100	98.5	69181	PASS
175	174	5	9	7.8	5418	PASS
176	174	95	101	96.8	66947	PASS
177	176	5	9	6.8	4552	PASS