

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062724\
 Data File : VY018719.D
 Acq On : 27 Jun 2024 15:09
 Operator : SY/MD
 Sample : P3077-07
 Misc : 7.04g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 1-S-1

Quant Time: Jun 28 03:15:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 14 07:08:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	92255	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	166679	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	153069	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	59043	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	44894	48.335	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.660%
35) Dibromofluoromethane	7.710	113	52091	51.232	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.460%
50) Toluene-d8	10.173	98	194252	49.679	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.360%
62) 4-Bromofluorobenzene	12.477	95	56530	45.784	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	91.560%
Target Compounds						
20) Methylene Chloride	4.692	84	15242	5.030	ug/l	94
64) Tetrachloroethene	10.715	164	2849	2.610	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062724\
Data File : VY018719.D
Acq On : 27 Jun 2024 15:09
Operator : SY/MD
Sample : P3077-07
Misc : 7.04g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
1-S-1

Quant Time: Jun 28 03:15:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 14 07:08:24 2024
Response via : Initial Calibration

