

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062821\
 Data File : VY005254.D
 Acq On : 28 Jun 2021 17:10
 Operator : SY/MD
 Sample : M2864-01
 Misc : 5.67G/5ML/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-206

Quant Time: Jun 29 01:18:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 07:57:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	21546	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	33224	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	32427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	16659	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	12742	71.264	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 142.520%	
35) Dibromofluoromethane	7.728	113	10624	54.221	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.440%	
50) Toluene-d8	10.185	98	38445	49.063	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 98.120%	
62) 4-Bromofluorobenzene	12.483	95	13241	50.710	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 101.420%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062821\
Data File : VY005254.D
Acq On : 28 Jun 2021 17:10
Operator : SY/MD
Sample : M2864-01
Misc : 5.67G/5ML/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-206

Quant Time: Jun 29 01:18:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 07:57:52 2021
Response via : Initial Calibration

