

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062821\
 Data File : VY005256.D
 Acq On : 28 Jun 2021 17:56
 Operator : SY/MD
 Sample : M2856-01
 Misc : 5.03G/5ML/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-01-9.5-10.0

Quant Time: Jun 29 01:19:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 07:57:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	7569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	11269	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	11225	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	5071	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	4362	69.446	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	138.900%
35) Dibromofluoromethane	7.722	113	3223	48.496	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.000%
50) Toluene-d8	10.185	98	13141	49.444	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.880%
62) 4-Bromofluorobenzene	12.483	95	4241	47.886	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062821\
Data File : VY005256.D
Acq On : 28 Jun 2021 17:56
Operator : SY/MD
Sample : M2856-01
Misc : 5.03G/5ML/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-01-9.5-10.0

Quant Time: Jun 29 01:19:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 07:57:52 2021
Response via : Initial Calibration

