

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062821\
 Data File : VY005259.D
 Acq On : 28 Jun 2021 19:06
 Operator : SY/MD
 Sample : M2889-01
 Misc : 6.25G/5ML/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-1(11.5-12)

Quant Time: Jun 29 01:19:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 07:57:52 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	212706	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	332807	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	322054	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	149238	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	116251	65.859	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 131.720%		
35) Dibromofluoromethane	7.728	113	102032	51.984	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 103.960%		
50) Toluene-d8	10.185	98	404905	51.586	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 103.180%		
62) 4-Bromofluorobenzene	12.483	95	135778	51.911	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 103.820%		
Target Compounds						
						Qvalue
16) Acetone	3.966	43	83568	204.937	ug/l	97
18) Methyl Acetate	4.497	43	2740	2.528	ug/l	95
25) 2-Butanone	7.002	43	8812	14.820	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062821\
Data File : VY005259.D
Acq On : 28 Jun 2021 19:06
Operator : SY/MD
Sample : M2889-01
Misc : 6.25G/5ML/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-1(11.5-12)

Quant Time: Jun 29 01:19:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 07:57:52 2021
Response via : Initial Calibration

