

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062823\
 Data File : VY014483.D
 Acq On : 28 Jun 2023 16:09
 Operator : KP/MD
 Sample : 03415-08
 Misc : 6.27g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ESII-TP1

Quant Time: Jun 29 02:19:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 17:18:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	161020	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	255242	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	248370	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	122161	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	116217	58.132	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.260%
35) Dibromofluoromethane	7.715	113	85218	54.040	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.080%
50) Toluene-d8	10.178	98	307347	49.030	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.060%
62) 4-Bromofluorobenzene	12.477	95	102887	45.975	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	91.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062823\
Data File : VY014483.D
Acq On : 28 Jun 2023 16:09
Operator : KP/MD
Sample : 03415-08
Misc : 6.27g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ESII-TP1

Quant Time: Jun 29 02:19:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 17:18:46 2023
Response via : Initial Calibration

