

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062823\
 Data File : VY014485.D
 Acq On : 28 Jun 2023 16:56
 Operator : KP/MD
 Sample : 03415-08RE
 Misc : 5.24g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ESII-TP1RE

Quant Time: Jun 29 02:19:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 17:18:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.782	168	163658	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	257924	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	246446	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	118732	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	113962	56.085	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 112.160%	
35) Dibromofluoromethane	7.709	113	85904	53.909	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.820%	
50) Toluene-d8	10.178	98	311527	49.180	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 98.360%	
62) 4-Bromofluorobenzene	12.477	95	98541	43.575	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 87.160%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062823\
Data File : VY014485.D
Acq On : 28 Jun 2023 16:56
Operator : KP/MD
Sample : 03415-08RE
Misc : 5.24g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ESII-TP1RE

Quant Time: Jun 29 02:19:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 17:18:46 2023
Response via : Initial Calibration

