



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY062823\  
Data File : VY014469.D  
Acq On : 28 Jun 2023 09:17  
Operator : KP/MD  
Sample : BFB  
Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y062323S.M  
Title : SW846 8260  
Last Update : Fri Jun 23 17:18:46 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 20.3         | 13480      | PASS                |
| 75             | 95              | 30              | 60              | 55.7         | 36955      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 66357      | PASS                |
| 96             | 95              | 5               | 9               | 6.6          | 4390       | PASS                |
| 173            | 174             | 0.00            | 2               | 1.7          | 1093       | PASS                |
| 174            | 95              | 50              | 100             | 99.3         | 65909      | PASS                |
| 175            | 174             | 5               | 9               | 8.1          | 5333       | PASS                |
| 176            | 174             | 95              | 101             | 96.2         | 63429      | PASS                |
| 177            | 176             | 5               | 9               | 6.9          | 4374       | PASS                |