

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY062920\
 Data File : VY003006.D
 Acq On : 29 Jun 2020 14:52
 Operator : SY/MD
 Sample : VY0629SBL01
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0629SBL01

Quant Time: Jun 30 06:15:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y062920S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 30 04:11:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	353857	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.69	114	572172	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	546194	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.42	152	278824	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.14	65	204951	58.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	116.64%	
35) Dibromofluoromethane	7.72	113	186252	52.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.36%	
50) Toluene-d8	10.18	98	700226	52.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.32%	
62) 4-Bromofluorobenzene	12.48	95	252989	54.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.10%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.72	84	48152	7.027	ug/l	93
52) Toluene	10.25	92	11667	1.228	ug/l	98
68) m/p-Xylenes	11.70	106	15753	2.148	ug/l	99
93) 1,2,4-Trichlorobenzene	15.01	180	10047	1.725	ug/l	95
94) Hexachlorobutadiene	15.11	225	3876	1.055	ug/l	96
95) Naphthalene	15.24	128	23444	2.204	ug/l	96
96) 1,2,3-Trichlorobenzene	15.42	180	9654	1.909	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY062920\
Data File : VY003006.D
Acq On : 29 Jun 2020 14:52
Operator : SY/MD
Sample : VY0629SBL01
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0629SBL01

Quant Time: Jun 30 06:15:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y062920S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 30 04:11:35 2020
Response via : Initial Calibration

