


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv062920\
Data File : VY002998.D
Acq On : 29 Jun 2020 10:11
Operator : SY/MD
Sample : BFB
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y062920S.M
Title : SW846 8260
Last Update : Tue Jun 30 04:11:35 2020



AutoFind: Scans 1761, 1762, 1763; Background Corrected with Scan 1752

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	16636	PASS
75	95	30	60	52.6	44666	PASS
95	95	100	100	100.0	84852	PASS
96	95	5	9	7.0	5905	PASS
173	174	0.00	2	1.5	1088	PASS
174	95	50	100	86.9	73770	PASS
175	174	5	9	7.9	5834	PASS
176	174	95	101	98.5	72642	PASS
177	176	5	9	7.8	5695	PASS