

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\vy062921\
 Data File : VY005279.D
 Acq On : 29 Jun 2021 16:41
 Operator : SY/MD
 Sample : M2864-01
 Misc : 6.22G/5ML/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-206

Quant Time: Jun 30 04:00:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 07:57:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	192305	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	313120	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	316126	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	157472	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	106111	66.492	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.980%
35) Dibromofluoromethane	7.728	113	100343	54.338	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.680%
50) Toluene-d8	10.185	98	385237	52.166	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.340%
62) 4-Bromofluorobenzene	12.483	95	133109	54.091	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\vy062921\
Data File : VY005279.D
Acq On : 29 Jun 2021 16:41
Operator : SY/MD
Sample : M2864-01
Misc : 6.22G/5ML/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-206

Quant Time: Jun 30 04:00:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 07:57:52 2021
Response via : Initial Calibration

