

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\vy062921\
 Data File : VY005284.D
 Acq On : 29 Jun 2021 18:37
 Operator : SY/MD
 Sample : M2885-01
 Misc : 5.86G/5ML/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 26644

Quant Time: Jun 30 13:48:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 07:57:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	225432	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	366336	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	353715	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	143753	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	118745	63.474	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.940%
35) Dibromofluoromethane	7.728	113	115281	53.359	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.720%
50) Toluene-d8	10.185	98	442846	51.255	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.520%
62) 4-Bromofluorobenzene	12.483	95	148364	51.532	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\vy062921\
Data File : VY005284.D
Acq On : 29 Jun 2021 18:37
Operator : SY/MD
Sample : M2885-01
Misc : 5.86G/5ML/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
26644

Quant Time: Jun 30 13:48:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 07:57:52 2021
Response via : Initial Calibration

