

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\vy062921\
 Data File : VY005290.D
 Acq On : 29 Jun 2021 20:56
 Operator : SY/MD
 Sample : M2895-03
 Misc : 6.14G/5ML/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-3

Quant Time: Jun 30 04:02:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 07:57:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	57	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.703	114	262	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	174	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.43
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/l	
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	10.179	98	277	44.828	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.660%
62) 4-Bromofluorobenzene	12.477	95	21	10.199	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	20.400%#
Target Compounds						
					Qvalue	
16) Acetone	3.954	43	26	237.935	ug/l #	40
67) Ethyl Benzene	11.715	91	40	6.726	ug/l #	42

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\vy062921\
Data File : VY005290.D
Acq On : 29 Jun 2021 20:56
Operator : SY/MD
Sample : M2895-03
Misc : 6.14G/5ML/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-3

Quant Time: Jun 30 04:02:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 07:57:52 2021
Response via : Initial Calibration

