



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\vy062921\
 Data File : VY005266.D
 Acq On : 29 Jun 2021 09:52
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062221S.M
 Title : SW846 8260
 Last Update : Wed Jun 23 07:57:52 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	22280	PASS
75	95	30	60	48.3	63931	PASS
95	95	100	100	100.0	132259	PASS
96	95	5	9	6.8	9052	PASS
173	174	0.00	2	1.1	1426	PASS
174	95	50	100	97.1	128435	PASS
175	174	5	9	7.7	9932	PASS
176	174	95	101	97.5	125285	PASS
177	176	5	9	7.7	9606	PASS