

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062922\
 Data File : VY009374.D
 Acq On : 29 Jun 2022 15:45
 Operator : KP/MD
 Sample : N3492-02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SOIL-PILE

Quant Time: Jun 29 17:44:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	150467	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	256913	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	246501	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	108813	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	68012	56.433	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.860%
35) Dibromofluoromethane	7.716	113	60558	42.828	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.660%
50) Toluene-d8	10.173	98	232813	53.702	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.400%
62) 4-Bromofluorobenzene	12.477	95	74117	36.047	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	72.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062922\
Data File : VY009374.D
Acq On : 29 Jun 2022 15:45
Operator : KP/MD
Sample : N3492-02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SOIL-PILE

Quant Time: Jun 29 17:44:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration

