

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062922\
 Data File : VY009377.D
 Acq On : 29 Jun 2022 16:55
 Operator : KP/MD
 Sample : N3516-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 R-1

Quant Time: Jun 30 05:12:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	141288	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	246227	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	234574	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	104679	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	64972	57.413	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.820%
35) Dibromofluoromethane	7.710	113	57625	42.522	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.040%
50) Toluene-d8	10.173	98	224648	54.067	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.140%
62) 4-Bromofluorobenzene	12.477	95	71169	36.116	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	72.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062922\
Data File : VY009377.D
Acq On : 29 Jun 2022 16:55
Operator : KP/MD
Sample : N3516-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
R-1

Quant Time: Jun 30 05:12:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration

