

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062922\
 Data File : VY009378.D
 Acq On : 29 Jun 2022 17:18
 Operator : KP/MD
 Sample : N3531-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 W-1

Quant Time: Jun 30 05:12:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 17:35:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	93733	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	159706	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	148253	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	64536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	40987	54.594	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.180%	
35) Dibromofluoromethane	7.710	113	37305	42.441	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 84.880%	
50) Toluene-d8	10.173	98	142719	52.958	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.920%	
62) 4-Bromofluorobenzene	12.477	95	42219	33.032	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 66.060%	
Target Compounds						
16) Acetone	3.942	43	4008	16.255	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062922\
Data File : VY009378.D
Acq On : 29 Jun 2022 17:18
Operator : KP/MD
Sample : N3531-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
W-1

Quant Time: Jun 30 05:12:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062822S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 17:35:34 2022
Response via : Initial Calibration

